

CORMAY IMMUNO-CONTROL II (L)

Nr kat. / Cat. No 4-290

Seria / Lot: 003-62

Data ważn. / Exp.: 2022.02

PARAMETR / COMPONENT	METODA / METHOD	WARTOŚĆ ŚRODKOWA / ASSIGNED VALUE	ZAKRES / RANGE
MIOGLOBINA MYOGLOBIN	TIA-wzmacniane lateksem Latex-enhanced TIA	98 ng/ml	83 – 113 ng/ml
FERRYTYNA FERRITIN	TIA-wzmacniane lateksem Latex-enhanced TIA	105 ng/ml	89 – 121 ng/ml
CALKOWITE IgE TOTAL IgE	TIA-wzmacniane lateksem Latex-enhanced TIA	70 IU/ml	53 – 88 IU/ml

CORMAY IMMUNO-CONTROL II (L)

Nr kat. / Cat. No 4-290

Seria / Lot: 003-62

Data ważn. / Exp.: 2022.02

PARAMETR / COMPONENT	METODA / METHOD	WARTOŚĆ ŚRODKOWA / ASSIGNED VALUE	ZAKRES / RANGE
MIOGLOBINA MYOGLOBIN	TIA-wzmacniane lateksem Latex-enhanced TIA	98 ng/ml	83 – 113 ng/ml
FERRYTYNA FERRITIN	TIA-wzmacniane lateksem Latex-enhanced TIA	105 ng/ml	89 – 121 ng/ml
CALKOWITE IgE TOTAL IgE	TIA-wzmacniane lateksem Latex-enhanced TIA	70 IU/ml	53 – 88 IU/ml

CORMAY IMMUNO-CONTROL II (H)

Nr kat. / Cat. No 4-290

Seria / Lot: 003-63

Data ważn. / Exp.: 2022.02

PARAMETR / COMPONENT	METODA / METHOD	WARTOŚĆ ŚRODKOWA / ASSIGNED VALUE	ZAKRES / RANGE
MIOGLOBINA MYOGLOBIN	TIA-wzmacniane lateksem Latex-enhanced TIA	312 ng/ml	281 – 343 ng/ml
FERRYTYNA FERRITIN	TIA-wzmacniane lateksem Latex-enhanced TIA	405 ng/ml	344 – 466 ng/ml
CALKOWITE IgE TOTAL IgE	TIA-wzmacniane lateksem Latex-enhanced TIA	405 IU/ml	344 – 466 IU/ml

CORMAY IMMUNO-CONTROL II (H)

Nr kat. / Cat. No 4-290

Seria / Lot: 003-63

Data ważn. / Exp.: 2022.02

PARAMETR / COMPONENT	METODA / METHOD	WARTOŚĆ ŚRODKOWA / ASSIGNED VALUE	ZAKRES / RANGE
MIOGLOBINA MYOGLOBIN	TIA-wzmacniane lateksem Latex-enhanced TIA	312 ng/ml	281 – 343 ng/ml
FERRYTYNA FERRITIN	TIA-wzmacniane lateksem Latex-enhanced TIA	405 ng/ml	344 – 466 ng/ml
CALKOWITE IgE TOTAL IgE	TIA-wzmacniane lateksem Latex-enhanced TIA	405 IU/ml	344 – 466 IU/ml