

CORMAY IMMUNO-MULTICAL

Cat. No: 4-287

Lot: 806-49

Exp.: 2020.02

PROTEIN	CALIBRATOR 1	CALIBRATOR 2	CALIBRATOR 3	CALIBRATOR 4	CALIBRATOR 5
ALBUMIN	6.60 g/l	12.60 g/l	24.90 g/l	50.50 g/l	103.40 g/l
$\alpha$ -1 ANTITRYPSIN	0.21 g/l	0.44 g/l	0.87 g/l	1.74 g/l	3.47 g/l
$\alpha$ -1 GLYCOPROTEIN ACID	0.13 g/l	0.25 g/l	0.51 g/l	1.03 g/l	2.06 g/l
ANTITHROMBIN III	0.038 g/l	0.076 g/l	0.175 g/l	0.362 g/l	0.724 g/l
COMPLEMENT C3	0.24 g/l	0.47 g/l	0.94 g/l	1.88 g/l	3.77 g/l
COMPLEMENT C4	0.041 g/l	0.082 g/l	0.164 g/l	0.327 g/l	0.654 g/l
CERULOPLASMIN*	0.095 g/l	0.165 g/l	0.303 g/l	0.596 g/l	1.220 g/l
HAPTOGLOBIN	0.21 g/l	0.40 g/l	0.82 g/l	1.63 g/l	3.28 g/l
IgA	0.33 g/l	0.65 g/l	1.31 g/l	2.61 g/l	5.24 g/l
IgG	1.55 g/l	3.10 g/l	6.21 g/l	12.40 g/l	24.82 g/l
IgM	0.14 g/l	0.30 g/l	0.59 g/l	1.16 g/l	2.34 g/l
TRANSFERRIN	0.40 g/l	0.79 g/l	1.58 g/l	3.16 g/l	6.32 g/l

\*Values compatible with CERULOPLASMIN reagents since LOT number xxx-xx27

CORMAY IMMUNO-MULTICAL

Cat. No: 4-287

Lot: 806-49

Exp.: 2020.02

PROTEIN	CALIBRATOR 1	CALIBRATOR 2	CALIBRATOR 3	CALIBRATOR 4	CALIBRATOR 5
ALBUMIN	6.60 g/l	12.60 g/l	24.90 g/l	50.50 g/l	103.40 g/l
$\alpha$ -1 ANTITRYPSIN	0.21 g/l	0.44 g/l	0.87 g/l	1.74 g/l	3.47 g/l
$\alpha$ -1 GLYCOPROTEIN ACID	0.13 g/l	0.25 g/l	0.51 g/l	1.03 g/l	2.06 g/l
ANTITHROMBIN III	0.038 g/l	0.076 g/l	0.175 g/l	0.362 g/l	0.724 g/l
COMPLEMENT C3	0.24 g/l	0.47 g/l	0.94 g/l	1.88 g/l	3.77 g/l
COMPLEMENT C4	0.041 g/l	0.082 g/l	0.164 g/l	0.327 g/l	0.654 g/l
CERULOPLASMIN*	0.095 g/l	0.165 g/l	0.303 g/l	0.596 g/l	1.220 g/l
HAPTOGLOBIN	0.21 g/l	0.40 g/l	0.82 g/l	1.63 g/l	3.28 g/l
IgA	0.33 g/l	0.65 g/l	1.31 g/l	2.61 g/l	5.24 g/l
IgG	1.55 g/l	3.10 g/l	6.21 g/l	12.40 g/l	24.82 g/l
IgM	0.14 g/l	0.30 g/l	0.59 g/l	1.16 g/l	2.34 g/l
TRANSFERRIN	0.40 g/l	0.79 g/l	1.58 g/l	3.16 g/l	6.32 g/l

\*Values compatible with CERULOPLASMIN reagents since LOT number xxx-xx27

CORMAY IMMUNO-MULTICAL

Nr kat.: 4-287

Seria: 806-49

Data ważn.: 2020.02

BIALKO	CALIBRATOR 1	CALIBRATOR 2	CALIBRATOR 3	CALIBRATOR 4	CALIBRATOR 5
ALBUMINA	6,60 g/l	12,60 g/l	24,90 g/l	50,50 g/l	103,40 g/l
$\alpha$ -1 ANTYTRYPSYNA	0,21 g/l	0,44 g/l	0,87 g/l	1,74 g/l	3,47 g/l
KWAŚNA $\alpha$ 1-GLIKOPROTEINA	0,13 g/l	0,25 g/l	0,51 g/l	1,03 g/l	2,06 g/l
ANTYTROMBINA III	0,038 g/l	0,076 g/l	0,175 g/l	0,362 g/l	0,724 g/l
SKŁADNIK C3 DOPEŁNIACZA	0,24 g/l	0,47 g/l	0,94 g/l	1,88 g/l	3,77 g/l
SKŁADNIK C4 DOPEŁNIACZA	0,041 g/l	0,082 g/l	0,164 g/l	0,327 g/l	0,654 g/l
CERULOPLAZMINA*	0,095 g/l	0,165 g/l	0,303 g/l	0,596 g/l	1,220 g/l
HAPTOGLOBINA	0,21 g/l	0,40 g/l	0,82 g/l	1,63 g/l	3,28 g/l
IgA	0,33 g/l	0,65 g/l	1,31 g/l	2,61 g/l	5,24 g/l
IgG	1,55 g/l	3,10 g/l	6,21 g/l	12,40 g/l	24,82 g/l
IgM	0,14 g/l	0,30 g/l	0,59 g/l	1,16 g/l	2,34 g/l
TRANSFERYNA	0,40 g/l	0,79 g/l	1,58 g/l	3,16 g/l	6,32 g/l

\*Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie od xxx-xx27

CORMAY IMMUNO-MULTICAL

Nr kat.: 4-287

Seria: 806-49

Data ważn.: 2020.02

BIALKO	CALIBRATOR 1	CALIBRATOR 2	CALIBRATOR 3	CALIBRATOR 4	CALIBRATOR 5
ALBUMINA	6,60 g/l	12,60 g/l	24,90 g/l	50,50 g/l	103,40 g/l
$\alpha$ -1 ANTYTRYPSYNA	0,21 g/l	0,44 g/l	0,87 g/l	1,74 g/l	3,47 g/l
KWAŚNA $\alpha$ 1-GLIKOPROTEINA	0,13 g/l	0,25 g/l	0,51 g/l	1,03 g/l	2,06 g/l
ANTYTROMBINA III	0,038 g/l	0,076 g/l	0,175 g/l	0,362 g/l	0,724 g/l
SKŁADNIK C3 DOPEŁNIACZA	0,24 g/l	0,47 g/l	0,94 g/l	1,88 g/l	3,77 g/l
SKŁADNIK C4 DOPEŁNIACZA	0,041 g/l	0,082 g/l	0,164 g/l	0,327 g/l	0,654 g/l
CERULOPLAZMINA*	0,095 g/l	0,165 g/l	0,303 g/l	0,596 g/l	1,220 g/l
HAPTOGLOBINA	0,21 g/l	0,40 g/l	0,82 g/l	1,63 g/l	3,28 g/l
IgA	0,33 g/l	0,65 g/l	1,31 g/l	2,61 g/l	5,24 g/l
IgG	1,55 g/l	3,10 g/l	6,21 g/l	12,40 g/l	24,82 g/l
IgM	0,14 g/l	0,30 g/l	0,59 g/l	1,16 g/l	2,34 g/l
TRANSFERYNA	0,40 g/l	0,79 g/l	1,58 g/l	3,16 g/l	6,32 g/l

\*Wartości kompatybilne z odczynnikami do oznaczania CERULOPLAZMINY serie od xxx-xx27